

Zhongshan Chengxiang Technology Co., Ltd.

No. 2 Zone B, Floor 3, Building F, Purong Science Park, No.1, Jinfeng Road, Jinbiao Village, Minzhong Town, Zhongshan City Guangdong Province China
Tel:0755-27835449 E-Mail:3614251723@qq.com

DECLARATION OF CONFORMITY

No.:TST20240302973-2EC

Applicant :Zhongshan Chengxiang Technology Co., Ltd.
Address :No.2 Zone B, Floor 3, Building F, Purong Science Park, No.1, Jinfeng Road, Jinbiao Village, Minzhong Town, Zhongshan City Guangdong Province China
Manufacturer :Zhongshan Chengxiang Technology Co., Ltd.
Address :No.2 Zone B, Floor 3, Building F, Purong Science Park, No.1, Jinfeng Road, Jinbiao Village, Minzhong Town, Zhongshan City Guangdong Province China
Product :Wood Moisture Meter
Model(s) :MT10, HY16B
Trademark :N/A

European Agency Information

Company

Name :VIKO PROCONSULT UK LTD
Address :Unit B2 Office New Smithfield Market, Whitworth Street East, Manchester, M11 2WJ, UK
City Manchester
Federal States UK (The United Kingdom/England)
Postal Code M11 2WJ

Email : vikouk@outlook.com
Tel : +44 2039363935

Test Standards:

EN IEC 61326-1:2021
EN IEC 61000-3-2:2019/A1:2021
EN 61000-3-3:2013/A1:2019/A2:2021

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

The certificate applies to the tested sample above mentioned only and shall not imply an assessment of the whole production. It is only valid in connection with the test report number:TST20240302973-2ER.



Xing hong Fan
(Manager)

May 16, 2024